



valid until: September 22, 2028

Fraunhofer

TESTED[®] DEVICE

Mettler-Toledo GmbH
PHD779lift

Report No. ME 2308-1443

Product series
Hygienic Design

Qualification Certificate

This is to certify that the product mentioned above, provided by

Mettler-Toledo GmbH
Greifensee, Switzerland

has been awarded a Fraunhofer certificate TESTED DEVICE
bearing the report number ME 2308-1443.

The PHD779lift series is principally suitable for use in hygienic
areas up to the following GMP Class according to EU GMP
Annex 1:

Suitability
up to GMP Class B

However, this only applies to the tested system in a resting state;
an overall assessment of the manufacturing environment would
need to be made after its installation.

The device may only be used in GMP Class B when it is certain
that the geometry of device does not impair the quality of the
laminar flow.

This document only
applies to the named
product in its original state
and is valid for a period of
5 years from the date the
first document was issued.
The document can be
verified under
www.tested-device.com.

Detailed information and
parameters of the test
environment can be found
in the Fraunhofer IPA test
report.

ME 2308-1443
Report No. first document

Stuttgart, September 22, 2023
Place, date of first document issued

--
Report No. current document

--
Place, current date

on behalf of 
Dr.-Ing. Frank Bürger, Project Manager Fraunhofer IPA