



Fraunhofer

**TESTED[®]
DEVICE**

Rite-Hite Corporation
LS Clean US 05/2021
Report No. RI 2108-1247

Statement of
Qualification

Single product
Biological Resistance

Statement of Qualification · Single product

Customer
 Rite-Hite Material Handling Equipment (Kunshan) Co., Ltd.
 110 Donglong Road
 Kunshan Jiangsu 215300
 China

Component tested

Category: Material
 Subcategory: Plastics
 Product name: VINYL,27OZ,WHITE,TOPCOAT, 118"
 (delivery date to Rite-Hite: 4/8/2021; color: 907907; article number: 8556)

Biological resistance test

Standards/Guidelines: ISO 846
 The norms stated generally refer to the version valid at the time of the tests.

Test environment parameters: Microbiological laboratory:.....S2

Test procedure parameters:

- Procedure A (resistance to fungi) using spore suspension of spores containing the following test strains:
 - *Aspergillus niger* ASM 1957 – *Penicillium pinophilum* ASM 1944
 - *Chaetomium globosum* ASM 1962 – *Trichoderma virens* ASM 1963
 - *Paecilomyces variotii* ASM 1961
- Procedure C (resistance to bacteria) using bacteria suspension containing the following test strain: *Pseudomonas aeruginosa* DSM 1253
- Incubation at 29±1 °C with a relative humidity of ≥95 %; visually inspection after four (4) weeks

Test result / Classification

The biological resistance of VINYL,27OZ,WHITE,TOPCOAT, 118" regarding to growth intensity was investigated in accordance with ISO 846 and classified with the following result:

Biological resistance	Growth intensity	Classification
Procedure A (resistance to fungi)	5	none
Procedure C (resistance to bacteria)	0	excellent
Overall result	none	

The classification is based on a worst-case consideration of the Procedures A and C. In the process, growth intensity was assessed according to the classification system used in ISO 846:

Classification: Procedure A (resistance to fungi)
 0 = excellent 2, 3 = weak
 1a, 1b, 1c = good 4, 5 = none

Classification: Procedure C (resistance to bacteria)
 0 = excellent 2 = weak
 1 = good 3 = none

The measuring devices used for the qualification tests are calibrated at regular intervals; their results can be traced back to national and international standards. In cases where no national standards exist, the test procedure implemented complies with the technical regulations and norms applicable at the time of the test. The relevant documentation can be viewed on request at any time.

Detailed information and parameters of the test environment can be found in the Fraunhofer IPA test report.