



Fraunhofer

TESTED[®] DEVICE

Roxtec International AB
LABSEAL C LTX
Report No. RO 2508-1666

Single product Riboflavin Test (Equipment)

Qualification Certificate

This is to certify that the product mentioned above, provided by

Roxtec International AB
Lyckeby, Sweden

has been awarded a Fraunhofer certificate TESTED DEVICE
bearing the report number RO 2508-1666.

The LABSEAL C LTX WITH COUNTER FRAME 0-40 (color: white)
can be cleaned good using a simple wiping procedure with
ultrapure water. However, the fluorescence test identified some
critical areas. These areas have to be cleaned especially thoroughly
or using a more complex procedure, e.g. by removing certain parts
before cleaning.

Product	Assessment of cleanability
LABSEAL C LTX WITH COUNTER FRAME 0-40	good

The ability to clean the LABSEAL C LTX WITH COUNTER FRAME
0-40 effectively could be improved by eliminating most of the
depressions or undercuts.

This document only
applies to the named
product in its original state
and is valid for a period of
5 years from the date the
first document was issued.
The document can be
verified under
www.tested-device.com.

Detailed information and
parameters of the test
environment can be found
in the Fraunhofer IPA test
report.

RO 2508-1666
Report No. first document

Stuttgart, September 24, 2025
Place, date of first document issued

--
Report No. current document

--
Place, current date

on behalf of 
Dr.-Ing. Frank Bürger, head of business unit Testing and Certification



Fraunhofer
IPA