



Fraunhofer

TESTED<sup>®</sup>  
DEVICE

CAITRON GmbH  
HSH - support arm system  
**Report No. CA 2505-1630**

DUPLICATE

Statement of  
Qualification

Product series  
Hygienic Design

Customer	CAITRON GmbH Gewerbepark Edelweiss 4 88138 Weissensberg Germany
Tested product	
Category:	Working Place and Operator
Subcategory:	Work Equipment
Product name:	HSH support arm system (manufacturing date: 2/26/2025; material: stainless steel) consisting of the following components: <ul style="list-style-type: none"><li>• HSH flange standard 101216</li><li>• HSH flange with adapter for keyboard holder 101224</li><li>• HSH cover with bayonet lock IP69 101218</li><li>• HSH I-holder (A528) 101227</li><li>• HSH L-bracket (A348, B251) 101228</li><li>• HSH L-bracket short (A85, B176) 101219</li><li>• HSH U-bracket (A346, B500, C251) 101359</li><li>• HSH U-bracket with two swivel joints (A346, B500, C277) 101360</li><li>• HSH standpipe (A1228) 101238</li><li>• HSH standpipe (A1228) with adapter 101314</li></ul>
Assessment of conformity to GMP regulations as well as to EHEDG conception and design recommendations	
Standards/guidelines:	EU GMP Annex 1; EHEDG Doc. 8; DIN EN 1672-2; ISO 14159 The norms stated generally refer to the version valid at the time of the tests.
Assessment criteria:	<ul style="list-style-type: none"><li>• Materials utilized</li><li>• Material pairings</li><li>• Installed components</li><li>• Geometries of components used</li><li>• Joining methods</li><li>• Detailed constructional solutions</li><li>• Manufacturing processes</li><li>• Surface coatings/coating systems</li></ul> <p>The suitability of the operating utility for use in a GMP-conform manufacturing environment is ascertained on the basis of the assessment of these criteria with the aid of expert knowledge. The assessment focuses mainly on the avoidance of contamination as well as on the ability to clean and disinfect the operating utility.</p>

Test result / Classification


The series HSH support arm system is principally suitable for use in hygienic areas up to the following GMP Class according to EU GMP Annex 1:

Suitability
up to GMP Class A

However, this recommendation only pertains to the operating utility when in a resting state. An overall assessment can only be made after its installation in the production line.

The measuring devices used for the qualification tests are calibrated at regular intervals; their results can be traced back to national and international standards. In cases where no national standards exist, the test procedure implemented complies with the technical regulations and norms applicable at the time of the test. The relevant documentation can be viewed on request at any time.

Detailed information and parameters of the test environment can be found in the Fraunhofer IPA test report.

Fraunhofer Institute for Manufacturing Engineering and Automation IPA	CA 2505-1630 Report No. first document	Stuttgart, August 15, 2025 Place, date of first document issued
Business unit Testing and Certification	-- Report No. current document	-- Place, current date
Nobelstrasse 12 70569 Stuttgart Germany	on behalf of Dr.-Ing. Frank Bürger, head of business unit Testing and Certification	

This document only applies to the named product in its original state and is valid for a period of 5 years from the date the first document was issued. The document can be verified under [www.tested-device.com](http://www.tested-device.com).