



valid until: August 29, 2030

Fraunhofer

TESTED[®] DEVICE

ads-tec Industrial IT
SHW9000

Report No. AD 2507-1652

Product series
Hygienic Design

Qualification Certificate

This is to certify that the product mentioned above, provided by

ads-tec Industrial IT GmbH
Nürtingen, Germany

has been awarded a Fraunhofer certificate TESTED DEVICE
bearing the report number AD 2507-1652.

The Smart Hygienic Panels series SHW9000 is principally suitable
for use in hygienic areas up to the following GMP Class according
to EU GMP Annex 1:

Suitability
up to GMP Class A

However, this only applies to the tested operating utility in a res-
ting state. An overall assessment of the manufacturing environ-
ment would need to be made after its installation.

This document only
applies to the named
product in its original state
and is valid for a period of
5 years from the date the
first document was issued.
The document can be
verified under
www.tested-device.com.

Detailed information and
parameters of the test
environment can be found
in the Fraunhofer IPA test
report.

AD 2507-1652
Report No. first document

Stuttgart, August 29, 2025
Place, date of first document issued

--
Report No. current document

--
Place, current date

on behalf of 
Dr.-Ing. Frank Bürger, head of business unit Testing and Certification



Fraunhofer
IPA