



valid until: August 14, 2030

Fraunhofer

TESTED[®] DEVICE

ISOONE
KILLBAC

Report No. LA 2411-1574

Single product
Chemical
Resistance

Qualification Certificate

This is to certify that the product mentioned above, provided by

ISOONE

Saint-Jean, France

has been awarded a Fraunhofer certificate TESTED DEVICE
bearing the report number LA 2411-1574.

The chemical resistance of the Coating KILLBAC (color: RAL 9003)
was classified according to ISO 4628-1 and VDI 2083 Part 17 with
the following result:

Chemical resistance (24 h)			
Formalin 37 %	excellent	Peracetic acid 15 %	excellent
Ammoniac 25 %	excellent	Hydrochloric acid 5 %	excellent
Hydrogen peroxide 30 %	excellent	Isopropanol 100 %	excellent
Sulfuric acid 5 %	excellent	Sodium hydroxide 5 %	excellent
Phosphoric acid 30 %	excellent	Sodium hypochlorite 5 %	excellent

This document only
applies to the named
product in its original state
and is valid for a period of
5 years from the date the
first document was issued.

The document can be
verified under
www.tested-device.com.

Detailed information and
parameters of the test
environment can be found
in the Fraunhofer IPA test
report.

LA 2411-1574
Report No. first document

Stuttgart, August 14, 2025
Place, date of first document issued

--
Report No. current document

--
Place, current date

on behalf of 
Dr.-Ing. Frank Bürger, head of business unit Testing and Certification



Fraunhofer
IPA