



Fraunhofer

TESTED[®] DEVICE

ISOONE
SIRIUS (SUR1084EPF)
Report No. LA 2411-1574

Single product Riboflavin Test (Equipment)

Qualification Certificate

This is to certify that the product mentioned above, provided by

ISOONE
Saint-Jean, France

has been awarded a Fraunhofer certificate TESTED DEVICE
bearing the report number LA 2411-1574.

The luminaire SIRIUS OPMI LED 450/450 5500/1 44/840 EPF (color: white) can be cleaned simply by wiping it with ultrapure water. However, the fluorescence test identified a few critical areas. These areas have to be cleaned especially thoroughly or using a more complex procedure, e.g. by removing certain parts before cleaning.

System component	Assessment of cleanability
SIRIUS OPMI LED 450/450 5500/1 44/840 EPF	good

This document only applies to the named product in its original state and is valid for a period of 5 years from the date the first document was issued. The document can be verified under www.tested-device.com.

Detailed information and parameters of the test environment can be found in the Fraunhofer IPA test report.

LA 2411-1574
Report No. first document

Stuttgart, August 14, 2025
Place, date of first document issued

--
Report No. current document

--
Place, current date

on behalf of 
Dr.-Ing. Frank Bürger, head of business unit Testing and Certification